

Measurements International Metrology is Our Science, Accuracy is Our Business[™]

1340A HIGH PRECISION VOLTAGE DIVIDER

Simplify Your Procedures, Simplify Your Work.



Featuring

- Stability
- Cost-effectiveness
- Performance-based Results
- No Self-Alignment Required

Overview

Never before has DC voltage calibration and/or verification of DVMs and calibrators been so easy or reliable. Measurements International's recently launched 1340A is the metrology industry's leading choice thanks to its user friendly design, backed with the best features for optimal performance.

Model 1340A is another fine example of MI's history and world-leading experience in resistance. We invite metrologists and calibration technicians in national, military and third-party calibration laboratories to compare the performance of the 1340A against any products on the market today.

Feature	Benefit		
10:1, 100:1 and 1000:1 reference divider outputs to 1100 V. Accuracy divides input voltage by 10, 100,	Extreme precision to compare direct voltage levels of various sources to a 10 V voltage reference		
1000 or 10000. Maximum input voltage is 1100 V.	standard like a 1330A, 732B or 732C.		
Industry leading specifications requires no Self-Alignment or Calibration prior to use.	Customers no longer need to self-align or calibrate prior to each use. Saves time, money and frustration!		
Utilizes a special design network of high precision	Shields divider resistors from outside noise and		
resistors mounted in a temperature-controlled	provides temperature stability to improve		
chamber	performance.		
Front panel direct connection to calibrator; divider output connection to DVM for testing; both done with supplied cables.	Ease of use, saves time and money.		
Internally mounted temperature sensor PT100	Users can connect to front panel and monitor internal oven.		
New special hand-selected resistors and configuration	Lengthy, self-alignment no longer required to create divider network.		
	Industry-leading advancement in the DC Voltage		
Calibration of Divider performed directly against a	Divider commercial products which delivers		
	exceptional performance.		



Measurements International Metrology is Our Science, Accuracy is Our Business[™]

1340A HIGH PRECISION VOLTAGE DIVIDER



An intuitive and easy to use instrument

Figure 1. Example: 1000 V in divided to 10 V out to be measured



Figure 2. Example: The above diagram illustrates using a 1330A, 732B or 732C Reference in the connection sequence. A DVM can be used as a NULL detector to determine the offset of the 57XX calibrator on the 10:1 and 100:1 ratio.



Figure 3. 24 Hour Stability of Temperature Chamber



Measurements International Metrology is Our Science, Accuracy is Our Business[™]

1340A HIGH PRECISION VOLTAGE DIVIDER

Specifications: Rev 5

Ratio Range		Uncertainty (ppm)				
		7 Day	30 Day	90 Day	1 Year	input voltage (v)
Stage 1	10:1	0.04	0.2	0.5	2.0	10 V
Stage 2	10:1	0.04	0.2	0.5	2.0	10 V
Stage 3	10:1	0.04	0.2	0.5	2.0	110 V
Stage 4	10:1	0.04	0.2	0.5	2.0	1100 V
100:1		0.1	0.3	1.0	3.0	
1000:1		0.1	0.3	1.0	3.0	
10000:1		0.1	0.4	1.0	4.0	

Internal Temperature Stability	± 0.1 °C over a 1 Year Period	
Ambient Temperature Range	23 °C ± 5 °C	
Initial Warm-up Time	24 Hours	
Ambient Humidity Range	20 % to 90 % Non-condensing	
Storage	-50 °C to +50 °C	
Alignment	Not Required	
Isolation to Earth	> 10 ¹² Ω	
Direct Cable to 57XX	Provided	
PT 100 Cable	Provided	
Warranty	Standard 2 Year Parts & Labour	

* These specifications are based on the drift of the unit over time.

* The specifications listed above must be used in conjunction with the absolute calibration of the ratio and its uncertainty budget.

Dimensions $(L \times W \times H)$: 436 x 449 x 133 (mm)

Weight: 9 kg

Shipping Weight: 13 kg

Mains Power: 85 to 264 V_{ac} , 47 to 440 Hz

Corporate Headquarters Measurements International PO Box 2359, 118 Commerce Drive

Prescott, Ontario, Canada K0E 1T0 Phone: 613-925-5934

Fax: 613-925-1195 Email: sales@mintl.com Toll Free: 1-800-324-4988

© Copyright 2022 Measurements International Limited. All rights reserved.

Form MI 66, Rev. 17, Dated 2022-12-01 (QAP19, App. "N")